



Silicone Facial Mask
Acute Dermal Irritation Study
FINAL REPORT

Client: Metalware Technology Corp.
Testing Institution: SGS Taiwan Ltd.
Report No. : UB/2013/40495
Report Date: 2013/05/13

1. The content of this report is invalid if it is not presented as the entire report.
2. Any unauthorized alteration, forgery or falsification of the content or appearance of this report is unlawful and offenders may be prosecuted to the fullest extent of the law.
3. The results shown in this test report refer only to the article(s) tested.

CONCLUSION

The results showed that there was no any local toxic effects in treatment group, and there was no mortality. Therefore, a single topical application of 0.5 g of “Silicone Facial Mask” did not cause skin irritation.

Table 1. Grade of clinical observation of each rabbit

Animal No.	Items for Grading	Clinical Observation (time point/h)			
		1	24	48	72
0404-SR-01	Erythema and eschar formation	0	0	0	0
	Oedema formation	0	0	0	0
0404-SR-02	Erythema and eschar formation	0	0	0	0
	Oedema formation	0	0	0	0
0404-SR-03	Erythema and eschar formation	0	0	0	0
	Oedema formation	0	0	0	0

Table 2. Bodyweight of each rabbit

Sex	Animal No.	Starting Day (kg)	Completion Day (kg)
Female	0404-SR-01	2.3	2.3
	0404-SR-02	2.2	2.3
	0404-SR-03	2.2	2.2